

**DECLARATION OF CONFORMITY**

Manufacturer: **Mercator Medical (Thailand) Ltd.**  
88/8 Moo.12 Tambon Kampaengphet,  
Amphur Rattaphum, Songkhla 90180 Thailand

Declare under its sole responsibility that non-sterile powder free nitrile protective gloves, diamond textured

The products described above are in conformity with the provision of the PPE Regulation (EU) 2016/425 as a Category III product, Type B where such is the case, with standards EN ISO 374-1:2016+A1:2018, EN ISO 374-2:2019, EN ISO 374-4:2019, EN ISO 374-5:2016, EN 16523-1:2015+A1:2018 and EN ISO 21420: 2020

The products are subject of EU Type Examination (Module B) and Module C2 on-going Conformity under certificate No. 2777/14031-07/E00-00 issued by notified body:  
**SATRA Technology Europe Limited (Notified Body number 2777)**  
Bracetown Business Park, Clonee Dublin 15, Dublin, Ireland

The products are subject of UKCA Type Examination (Module B) and Module C2 On-going Conformity under certificate No. AB0321/17234-02/E00-00 issued by approved body:  
**SATRA Technology Centre Limited (Approved Body number 0321)**  
Wyndham Way, Telford Way, Kettering. Northamptonshire. NN16 8SD. United Kingdom

This declaration is supported by the Quality System approval to ISO13485:2016/EN ISO13485:2016 and ISO 9001:2015 issued by:  
**SGS United Kingdom Ltd**  
Rossmore Business Park Ellesmere Port Cheshire CH65 3EN UK

Date and place of issue:  
12.12.2022, Thailand

Signed on the behalf of the Manufacturer:



Mr. Praneet Inthajak  
Senior QA/RA Manager  
Mercator Medical (Thailand) Ltd.

**Mercator Medical (Thailand) LTD.**

88/8 Moo 12  
Tambon Kampaengphet  
Amphur Rattaphum  
Songkhla 90180 Thailand

ph.: +66 74 584 222  
fax.: +66 74 584 223  
enquiry@mercator.co.th  
www.mercator.co.th

**Sales Office**  
**Mercator Medical S.A.**  
ul. H. Modrzejewskiej 30  
31-327 Kraków, Poland

ph.: +48 (12) 66 55 480  
fax.: +48 (12) 66 55 415  
medical@mercatormedical.eu  
www.mercatormedical.eu