

# User instruction

The instruction below should be used in conjunction with detailed information on the packaging.

## Short description of the product

Nitrile examination, powder-free, non-sterile for disposable use. Product Reference "nitylex® black".

## Full description of the product

Raw material	: nitrile
External surface	: textured
Internal surface	: polymerized + chlorinated
Cuff	: beaded
Colour	: black
Shape	: ambidextrous, fitting to the right and left hand
Size range	: XS (5-6), S (6-7), M (7-8), L (8-9), XL (9-10)
AQL	: 1.5 or customer required
Quantity in packaging	: 100 pcs. by weight
Shelf life	: 3 years (from the date of manufacturing)

## Storage instructions

It is recommended to store the gloves in dry place, in the temperature of 5-35°C and to protect them against direct sunlight and fluorescent light.

Keep the gloves in a distance of not less than 1m from heating devices, sources of fire and ozone.

Do not keep in direct vicinity of solvents, oils, fuels and lubricants.

## Food contact

Gloves are marked with food contact symbol  and comply with the requirements of European Regulation (EC) No 1935/2004 on materials, European Commission Directive 93/11/EEC and Council of Europe Resolution AP (2004) 4 and with requirement of Regulation (EC) No 2023/2006 on Good Manufacturing Practice. Gloves are suitable for handling any type of food and have been tested for Overall Migration Test acc. EN 1186:

Extraction conditions (tested for 0.5 h in 40°C)	Test Result (limit < 10 mg/dm <sup>2</sup> )
3% Acetic acid	Pass
10% Ethanol	Pass
20% Ethanol	Pass
50% Ethanol	Pass
95% Ethanol	Pass
Isooctane	Pass

## MDR classification & compliance

Gloves are classified as class I Medical Device as per MDR 2017/745 and comply to standards: EN 455-1:2020, EN 455-2:2015, EN 455-3:2015, EN 455-4:2009, EN ISO 15223-1:2016, EN ISO 20417:2021

## PPE classification & compliance

Gloves are category III Personal Protective Equipment as per Annex II of the Regulation 2016/425 and comply to standards: EN ISO 21420: 2020, EN ISO 374-1:2016+A1:2018 (Type B), EN ISO 374-2:2019, EN 16523-1:2015, EN ISO 374-4:2019, EN ISO 374-5:2016.

EU Type Examination Certificate issued by: SATRA (Notified Body No. 2777)

UKCA Type Examination Certificate issued by: SATRA (Approved Body No. 0321)

Checking of PPE manufactured:

CE 2777

SATRA Technology Europe Limited.  
Bracetown Business Park.  
Clonee.  
D15YN2P.  
Republic of Ireland.

UK  
CA 0321  
SATRA Technology Centre Limited.  
Wyndham Way, Telford Way,  
Kettering, Northamptonshire,  
NN16 8SD,  
United Kingdom.

Declaration of Conformity and Instruction of Use with the information about the importer are available at: [www.mercatormedical.eu](http://www.mercatormedical.eu)

## Intended use

These are non-sterile examination and protective gloves for single use, intended for use in medical field to: protect patient and user from cross-contamination, conducting medical examinations, diagnostic and therapeutic procedures and for handling medical contaminated material. Gloves are classified as Medical Devices Class I and as a Personal Protective Equipment category III. Their design and labelling correspond to the requirements of the European Medical Device Regulation 2017/745 and the European Regulation 2016/425 on Personal Protective Equipment and included the PPE Regulation (EU) 2016/425 as brought into UK law and amended as a Category III product. Gloves should be used solely according to their intended application.

## Precautions and indications for use

Dry hands before putting the gloves on. Before usage, inspect the gloves for any defect or imperfections. Use at least 1 pair of gloves for one patient and one procedure, these are disposable gloves. Do not let chemical substances get under the gloves through the cuff. If a chemical substance reaches the skin, wash it away immediately with plenty of water with soap. If the gloves get punctured, torn or broken during their use, take them off and put on the new ones. Avoid using gloves dirty in the inside as they may cause irritation leading to skin inflammation or more serious damages. The gloves should not be used in contact with open fire and to protect against any sharp tools. The gloves are not intended for welding, electric shock protection, ionizing radiation or from the effect of hot or cold objects.

This information does not reflect the actual duration of protection in the workplace and the between mixtures and pure chemicals.

Degradation results indicate the change in puncture resistance of the gloves after exposure to the challenge chemical. The chemical penetration resistance has been assessed under laboratory conditions from samples taken from the palm only (except in case where glove is equal to or over 400 mm – where the cuff is tested also) and relates only to the chemical tested and to the tested specimen. It can be different if the chemical is used in a mixture.

The penetration resistance has been assessed under laboratory conditions and relates only to the tested specimen. It is recommended to check that the gloves are suitable for the intended use because the conditions at the workplace may differ from the type test depending on the temperature, abrasion and degradation.

When used, protective gloves may provide less resistance to the dangerous chemical due to changes in physical properties. Movements, snagging, rubbing, degradation caused by the chemical contact etc. may reduce the actual use time significantly. For corrosive chemicals, degradation can be the most important factor to consider in selection of chemical resistant gloves.

Gloves are suitable for special purposes as they are examination gloves where risk of wrist injury caused by chemicals is considered to be minimal. Length suitable for tasks that require hand protection. Glove minimum length in accordance to EN 455-2 standard.

## Components / hazardous components

Some gloves may contain components known to be a possible cause of allergy for person allergic to them, who may develop contact irritation and/or allergic reaction. In case of an allergic reaction, seek medical assistance immediately.

## Disposal

Used gloves can be contaminated with contagious or other hazardous substances. They should be disposed of in accordance with local regulation. Gloves should be buried or burned under controlled conditions.

## Manufacturer

MERCATOR MEDICAL (Thailand) Ltd.  
88/8 Moo 12 Tambon Kampaengphet,  
Amphur Rattaphum, Songkhla 90180, Thailand.  
[www.mercator.co.th](http://www.mercator.co.th)

## Authorized Representative/Importer

Mercator Medical S.A.  
Address: ul. H. Modrzewskiej 30  
31-327 Krakow, Poland

Permeation performance levels as per EN ISO 374-1:2016			EN ISO 374-4:2019 Degradation [%]
Level1>10min	Level2>30 min	Level3>60 min, level4>120min, Level5>240min, Level6>480m in	
Test results acc. to EN 16623-1:2015			
Chemical	Level		
35% Ethanol	6		55.0
40% Isopropanol	6		68.7
10% Acetic acid	4		53.5
50% Benzalkonium chloride*	6		29.5
4% Chlorhexidine digluconate **	6		32.9
10% Phosphoric acid	6		14.0
40% Sodium hydroxide (K)	6		2.6
12% Sodium hypochlorite	6		22.7
50% Sulphuric Acid	6		21.1
5% Ethidium Bromide	6		32.9
3% Hydrogen peroxide	6		44.0
30% Hydrogen peroxide (P)	2		52.8
37% Formaldehyde (T)	5		20.0
50% Glutaraldehyde	6		22.9
0.1% Phenol	6		24.7






















EN ISO 374-4 : Degradation results indicate the change in puncture resistance of the gloves after exposure to the challenge chemical.

Noted : \* Permeation rate 5 µg/cm<sup>2</sup>/min.


\*\* Permeation rate 7 µg/cm<sup>2</sup>/min.

Test acc. To EN ISO 374-2:2019 – Level 2 (ISO 2859)		Test acc. To EN ISO 374-3:2016	
Performance level	AQL	Protection against bacteria & fungi	Pass
Level 3	< 0.65	Protection against viruses	Pass
Level 2	< 1.5		
Level 1	< 4.0		

## Symbols used on the packaging

	Do not re-use / gloves are intended for single use		Non-sterile gloves		Powderfree gloves
	Do not use, if package is damaged		Keep away from solar and fluorescent light		Keep away from ozone
	Keep away from moisture, store in a dry place		Temperature limitation / gloves store in temperature 5-35°C		100 gloves by weight
	Catalogue number		Gloves protecting against chemical dangers with digit literal codes		Package made from paper, qualify for recycling
	EC REP symbol should be used with name and address of EC representative		Date of manufacture		Package is treated as municipal waste
	Protection against bacteria, fungi and viruses		Food contact symbol (article is suitable for food contact, for details check the instruction for use)		Lot / batch number
	Manufacturer, symbol should be accompanied by name and address of Manufacturer		Consult instructions for use		
	Expiry date				

## HOW TO PUT THE GLOVES ON?



## HOW TO TAKE THE GLOVES OFF?

